


Study Title: Refitting Biofinity Multifocal Contact Lenses Wearers with MyDay Multifocal Contact Lenses.

Participants



Baseline Characteristics

N	66
Gender	
Female	52 (78.8%)
Male	14 (21.2%)
Age	
Mean	53.9
Median	53.0
Std. Dev	6.7
Minimum	41.0
Maximum	71.0
25%	48.8
50%	53.0
75%	58.0

1. Primary Outcome Measure Study Contact Lens Handling characteristics (Insertion & removal)

Table 1 - Study Contact Lenses Handling Characteristics

	EASE OF INSERTION	EASE OF REMOVAL	OVERALL LENS HANDLING
BIOFINITY			
N	66	66	66
Mean	90.7	88.3	89.1
Median	95.5	94.5	96.0
MYDAY			
N	66	66	66
Mean	84.8	91.5	82.4
Median	94.5	96.5	90.0

2- Secondary Outcome Measures LogMAR visual acuity (Mean of distance, intermediate and near)

Table 2 - Study Contact Lenses Overall HLHC VA (LogMAR)

	Overall HLHC VA (LogMAR)		
	Dom	Non-Dom	Bino
BIOFINITY			
N	66	66	66
Mean	0.01	0.01	-0.05
Median	0.00	0.01	-0.06
MYDAY			
N	66	66	66
Mean	0.01	0.00	-0.05
Median	0.01	0.00	-0.07

3. Adverse Events:

Ten adverse events presented included nine non-ocular and one ocular events. All adverse events were non-serious and did not invalidate the data collected. The ocular adverse event was non-device related.