

Evaluation of a new adhesive system for orthodontic braces for a 6-month period

Submission date 04/10/2020	Recruitment status No longer recruiting	<input type="checkbox"/> Prospectively registered <input type="checkbox"/> Protocol
Registration date 07/10/2020	Overall study status Completed	<input type="checkbox"/> Statistical analysis plan <input checked="" type="checkbox"/> Results
Last Edited 23/06/2023	Condition category Oral Health	<input type="checkbox"/> Individual participant data

Plain English summary of protocol

Background and study aims

Bracket failure (bracket detachment from the tooth) hinders the smooth progress of orthodontic treatment. Thus, a bracket failure rate that is as low as possible is of utmost importance for the clinician as well as for the patient. Recently, a new adhesive system was introduced. This system reduces the steps during bracket placement. Thus, the aim of this study is to evaluate the clinical performance (i.e. bracket failure as well as chair time) of this newly introduced adhesive pre-coated bracket system (PS) compared with a conventional bracket system (CS).

Who can participate?

Patients undergoing orthodontic treatment without tooth extraction

What does the study involve?

Participants receive orthodontic treatment where the mouth is divided into four segments that are randomly assigned to either the PS or CS bracket system. Bracket failure is measured over 6 months.

What are the possible benefits and risks of participating?

The risks of this study are the same as the risks of any routine non-extraction orthodontic treatment. No extra risks are present. The study will give insight concerning the chair time, i.e. the amount of time a patient spends in the chair for treatment, of these systems. Chair time is of great importance, particularly during the COVID-19 pandemic. Furthermore, this study will provide data on whether or not this system (PS) is as reliable as CS.

Where is the study run from?

Ondokuz Mayıs University (Turkey)

When is the study starting and how long is it expected to run for?

June 2018 to June 2019

Who is funding the study?

Investigator initiated and funded

Who is the main contact?
Dr Selma Elekdag-Türk
s.elekdag@omu.edu.tr

Contact information

Type(s)
Scientific

Contact name
Prof Selma Elekdag-Turk

ORCID ID
<https://orcid.org/0000-0002-2799-6501>

Contact details
Ondokuz Mayıs University
Faculty of Dentistry
Department of Orthodontics
Samsun
Türkiye
55200
+90 (0)5446403424
s.elekdag@omu.edu.tr

Additional identifiers

Clinical Trials Information System (CTIS)
Nil known

ClinicalTrials.gov (NCT)
Nil known

Protocol serial number
OMÜ KAEK 2018/416

Study information

Scientific Title
Evaluation of the flash-free adhesive system for a 6-month period: a split-mouth trial

Study objectives

The aim of this study was the evaluation of the following parameters:

1. The comparison of the bond failure and survival rates of the APC FF Adhesive Coated Appliance System (3M Unitek) with an OC system (Transbond XT Light Cure Adhesive Paste; 3M Unitek) for a 6-month period.
2. Comparison of the bond failure and survival rates of the upper and lower arches.
3. Comparison of the bond failure and survival rates for incisor, canine and premolar teeth.
4. Comparison of bracket failure and survival rates with respect to gender.

5. Comparison of the ARI scores upon bracket failure.
 6. Comparison of the chair time required for each bonding procedure.
- The null hypothesis was that there would be no difference in the aforementioned parameters.

Ethics approval required

Old ethics approval format

Ethics approval(s)

Approved 13/09/2018, Research Ethics Committee at the University of Ondokuz Mayıs (Former Dental School Building, Ondokuz Mayıs University, 55200 Atakum/Samsun, Turkey; +90 (0)362 312 19 19/2782; omutaek@gmail.com), ref: OMÜ KAEK 2018/416

Study design

Single-center randomized study

Primary study design

Interventional

Study type(s)

Treatment

Health condition(s) or problem(s) studied

Orthodontic malocclusion

Interventions

Patients willing to participate in this study receive orthodontic treatment with a split-mouth design, i.e. the mouth is divided into four experimental segments that are randomly assigned to either the pre-coated bracket system (PS) or a conventional bracket system (CS). The observation period for his study is 6 months.

Intervention Type

Device

Phase

Not Applicable

Drug/device/biological/vaccine name(s)

Adhesive pre-coated bracket system, operator-coated conventional bracket system

Primary outcome(s)

Bracket failure, i.e. bracket detachment from the tooth, is visually observed. Bracket failure can occur at any timepoint during the observation period. Bracket failure was recorded at the routine monthly controls of the patients.

Key secondary outcome(s)

Type of failure: the bracket may have detached from the tooth intact with adhesive or without adhesive. Furthermore, the bracket may have fractured. These types of failures can occur at any timepoint during the observation period. These types of failures were recorded at the routine monthly controls of the patients.

Completion date

18/06/2019

Eligibility

Key inclusion criteria

1. Willingness to participate in this research project
2. Fully erupted maxillary and mandibular teeth with intact buccal enamel
3. Angle Class I and mild Class II malocclusions
4. Non-extraction patients
5. Normal overbite (1-2 mm)
6. Absence of severe rotations. Severe rotations would have prevented bracket placement or correct bracket placement at the first appointment
7. Absence of pretreatment of teeth with any chemical agents
8. Good oral hygiene, i.e. no bleeding upon probing

Participant type(s)

Patient

Healthy volunteers allowed

No

Age group

Mixed

Sex

All

Total final enrolment

30

Key exclusion criteria

1. Patients with skeletal problems
2. Patients with missing teeth
3. Patients with systemic disease
4. Patients with a previous history of orthodontic treatment

Date of first enrolment

17/09/2018

Date of final enrolment

18/12/2018

Locations

Countries of recruitment

Türkiye

Study participating centre

Ondokuz Mayıs University
Faculty of Dentistry
Department of Orthodontics
Samsun
Türkiye
55210

Sponsor information

Organisation

Ondokuz Mayıs University

ROR

<https://ror.org/028k5qw24>

Funder(s)

Funder type

Other

Funder Name

Investigator initiated and funded

Results and Publications

Individual participant data (IPD) sharing plan

The datasets generated during and/or analysed during the current study are/will be available upon request from Prof. Dr Tamer Türk (turkset@superonline.com, tamert@omu.edu.tr) and Prof. Dr Selma Elekdag-Türk (elekdagturk@yahoo.com).

IPD sharing plan summary

Available on request

Study outputs

Output type	Details	Date created	Date added	Peer reviewed?	Patient-facing?
Results article		22/06/2023	23/06/2023	Yes	No