

# A comparison of widely used clinical contrast sensitivity tests: the relation between defocus specific contrast sensitivity and higher order aberrations

<b>Submission date</b> 01/12/2006	<b>Recruitment status</b> No longer recruiting	<input type="checkbox"/> Prospectively registered <input type="checkbox"/> Protocol
<b>Registration date</b> 01/12/2006	<b>Overall study status</b> Completed	<input type="checkbox"/> Statistical analysis plan <input checked="" type="checkbox"/> Results
<b>Last Edited</b> 06/01/2021	<b>Condition category</b> Other	<input type="checkbox"/> Individual participant data

## Plain English summary of protocol

Not provided at time of registration

## Contact information

### Type(s)

Scientific

### Contact name

Ms K W van Gaalen

### Contact details

University Medical Center  
Groningen  
Department of Ophtalmology  
P.O. Box 30001  
Groningen  
Netherlands  
9700 RB  
+31 (0)50 3619692  
k.van.gaaen@ohk.umcg.nl

## Additional identifiers

EudraCT/CTIS number

IRAS number

**ClinicalTrials.gov number**

**Secondary identifying numbers**

NL799, NTR812

## **Study information**

### **Scientific Title**

A comparison of widely used clinical contrast sensitivity tests: the relation between defocus specific contrast sensitivity and higher order aberrations

### **Acronym**

Defocus specific contrast sensitivity and spherical aberration

### **Study objectives**

Higher order aberrations, like spherical aberration, decreases visual performance.

### **Ethics approval required**

Old ethics approval format

### **Ethics approval(s)**

Medisch Ethische Toetsingscommissie, University Medical Center Groningen, date of MEC approval: 14 Oct 2005 (reference number: METc2005.188).

### **Study design**

Randomised parallel armed clinical trial

### **Primary study design**

Interventional

### **Secondary study design**

Randomised controlled trial

### **Study setting(s)**

Not specified

### **Study type(s)**

Treatment

### **Participant information sheet**

### **Health condition(s) or problem(s) studied**

No condition, healthy person

### **Interventions**

Best corrected visual acuity was determined with an Early Treatment Diabetic Retinopathy Study (ETDRS ) chart and the Spherical Aberration (SA) was measured with a wavefront analyzer (WASCA version 1.26.3, Asclepion Meditec, Jena, Germany).

The contrast sensitivity is measured with two computerised tests:

1. One with vertical sine-wave gratings (1.5-12 cpd) generated on a CRT (Cambridge Research Systems, Rochester, UK; Von Békésy tracking method)
2. The Holladay sine-wave (1.5 -18 cpd) modulated circular lines (HACSS) (M&S Technologies, Skokie, Illinois, USA),

and with six contrast sensitivity chart tests:

1. Pelli Robson contrast sensitivity test
2. Low contrast ETDRS-like optotype chart 2.5%
3. Low contrast ETDRS-like optotype chart 10%
4. Edge contrast sensitivity test: GECKO
5. Edge contrast sensitivity test: GECKO-100
6. Vector Vision

Contrast sensitivity is measured in mesopic (3 cd/m<sup>2</sup>) and photopic (160 cd/m<sup>2</sup>) conditions, using only the dominant eye. Tests were performed at optimal refractive state of the eye and at a variety of defocus situations (-2D to 2D).

### **Intervention Type**

Other

### **Phase**

Not Specified

### **Primary outcome measure**

Selection of the contrast sensitivity test which predicts the spherical aberration most reliably.

### **Secondary outcome measures**

1. Spherical aberration as function of age
2. RMS as function of age
3. Contrast sensitivity as function of age
4. Influence of defocus on contrast sensitivity

### **Overall study start date**

01/07/2005

### **Completion date**

30/06/2006

## **Eligibility**

### **Key inclusion criteria**

No ocular pathology

### **Participant type(s)**

Patient

### **Age group**

Not Specified

**Sex**

Not Specified

**Target number of participants**

48

**Total final enrolment**

48

**Key exclusion criteria**

1. Refractive correction larger than +/- 2 D
2. Cylindrical correction larger than 1.5 D
3. Cylindrical axis more then 20° from the horizontal or vertical axis

**Date of first enrolment**

01/07/2005

**Date of final enrolment**

30/06/2006

**Locations****Countries of recruitment**

Netherlands

**Study participating centre**

**University Medical Center**

Groningen

Netherlands

9700 RB

**Sponsor information****Organisation**

University Medical Center Groningen (The Netherlands)

**Sponsor details**

Department of Ophthalmology

P.O. Box 30001

Groningen

Netherlands

9700 RB

**Sponsor type**

Hospital/treatment centre

**Website**

<http://www.umcg.nl/azg/nl/english/azg/>

**ROR**

<https://ror.org/03cv38k47>

## Funder(s)

**Funder type**

Government

**Funder Name**

SenterNovem (The Netherlands)

## Results and Publications

**Publication and dissemination plan**

Not provided at time of registration

**Intention to publish date****Individual participant data (IPD) sharing plan****IPD sharing plan summary**

Not provided at time of registration

**Study outputs**

Output type	Details	Date created	Date added	Peer reviewed?	Patient-facing?
<a href="#">Abstract results</a>	results presented at ARVO	01/05/2007	06/01/2021	No	No